

EC Declaration of Conformity

Manufacturer/ Supplier Information:	BioFire Diagnostics, LLC 515 Colorow Drive Salt Lake City, Utah 84108, USA Phone: 1-801-736-6354 regulatory@BioFireDX.com http://www.BioFireDX.com
--	---

We, BioFire Diagnostics, LLC, declare under our sole responsibility, that the product

FilmArray® Respiratory Panel (RFIT-ASY-0124, RFIT-ASY-0125)

meets the provisions of the European Directive 98/79/EC for *in vitro* Diagnostic Medical Devices. The device is classified as an In Vitro Diagnostic (IVD) Device under Annex II list B. BioFire Diagnostics' quality system is registered to ISO 13485:2016 and EN ISO 13485:2016.

The following relevant standards have been met:

ISO 13485:2016/EN ISO 13485:2016 Medical devices – Quality Management System – Requirements for regulatory purposes
EN ISO 14971:2012 Medical devices – Application of risk management to medical devices
EN 13641:2002 Elimination or reduction of risk of infection related to in vitro diagnostic reagents
EN 62366:2008 Medical devices-Application of usability engineering to medical devices
EN 13612:2002 Performance evaluation of in vitro diagnostic medical devices
EN ISO 23640:2015 In vitro diagnostic medical devices – Evaluation of stability of in vitro diagnostic reagents
EN ISO 15223-1:2016 Medical Devices – Symbols to be used with medical device labels, labeling and information to be supplied – Part 1: General requirements
EN ISO 18113-1:2011 In vitro diagnostic medical devices – Information supplied by the manufacturer (labeling) – Part 1: Terms, definition and general requirements
EN ISO 18113-2:2011 In vitro diagnostic medical devices – Information supplied by the manufacturer (labeling) – Part 2: In vitro diagnostic reagents for professional use

Technical documentation demonstrating compliance as described in Annex IV of the European Directive 98/79/EC is kept by the manufacturer and can be made available by the authorized representative in Europe - QARAD bvba, Cipalstraat 3, B-2440 Geel, Belgium.

The notified body for this product is BSI (Notified body #0086; Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes, MK5 8PP, United Kingdom).

Salt Lake City, UT, USA 12-19-2018
 (Place and date of issue)



Randy Rasmussen
 Chief Executive Officer



515 Colorow Drive, Salt Lake City, UT 84108
 phone 1-801-736-6354 | fax 1-801-588-0507
www.BioFireDX.com